Swab Type Comparison Chart





Part No.	Material	Head Width	Head Thickness	Head Length	Handle Material	Breakpoint	Total Length
901101	Nylon fibre	50mm	1.3mm	20mm	polypropylene	80mm	154mm

Certificates: CE Marked, FDA approved Sterility: Gamma irradiated

Swab Image: Swab <t

Part No.	Material	Head Width	Head Thickness	Head Length	Handle Material	Breakpoint	Total Length
CY-98000	Nylon fibre	20mm	1.3mm	20mm	polypropylene	30mm	145mm

Certificates: CE Marked, FDA approved Sterility: Gamma irradiated




Part No.	Material	Head Width	Head Thickness	Head Length	Handle Material	Breakpoint	Total Length
CM-FS915	Nylon fibre	50mm	50mm	20mm	polypropylene	80mm	154mm

Certificates: CE Marked, FDA approved Sterility: Gamma irradiated



www.norgenbiotek.com





Certificates: CE marked, FDA approved Sterility: Ethylene Oxide Sterilization



www.norgenbiotek.com